

IMPARTIALITY SERVICE INNOVATION EFFICIENCY



## 3Phase Smart Power Meter Calibration Report


Report number : EC20140157

Issued date : 2014/07/24

Laboratory : Electrical and Temperature Calibration Laboratory

Address: No.6-6, Ronggong S Rd., Guanyin Township, Taoyuan County 328, Taiwan, R.O.C.

Laboratory Accreditation Number : 0061



### Taiwan Electric Research & Testing Center

Address: No.6-6, Ronggong S Rd., Guanyin Township, Taoyuan County 328, Taiwan, R.O.C.

TEL: 886-3-483-9090

FAX: 886-3-483-8119

E-mail: [customer\\_service@ms.tertec.org.tw](mailto:customer_service@ms.tertec.org.tw)

Website: [www.tertec.org.tw](http://www.tertec.org.tw)

- ◆ This report is calibration report.
- ◆ The testing result is only responsible to the tested sample.
- ◆ The report can be fully duplicated only. To excerpt any part of this report is invalid unless permitted by TERTEC.
- ◆ The contents on the report can not be used for advertisement, publication and merchandised activities.
- ◆ The report is invalid if without the seal on each page.
- ◆ Inquiry telephone: 886-3-483-9090 ext. 8201



財團法人

台灣大電力研究試驗中心  
Taiwan Electric Research & Testing Center

Electrical and Temperature Calibration Laboratory  
Report No. : EC20140157



## Calibration Report

Applicant : ICP DAS Co.,Ltd

Issued Date : 2014/07/24

Address : No. 111, Guangfu N. Rd., Hukou Township,  
Hsinchu County 30351, Taiwan, R.O.C.

Calibration Date : 2014/07/16

Equipment : 3Phase Smart Power Meter

Temperature : 23±2°C

Manufacturer : ICP DAS Co.,Ltd

Humidity : 50±10%

Model No. : PM-3133-100

Procedure Used : 60I-07-1812

Serial No. : PM3133ACR0A0FCP00001

## Remark :

1. Calibration items with the same manufacturing number of the external CT (1, 2, 3, ).
2. Calibration items RS485 communication interface and manufacturer of computer software (ICP DAS PM-3133-100, Version: V1.5) reader to indication value.

## Calibration Standard

Equipment	Manufacturer / Model No.	I.D. Number	Cal. Source / Cal. Date/ Report No/Cycle
Three Phase Standard	RADIAN/RD-30-231	300130	TERTEC / 103.03.25 / EC1030019 / 1 year

## 一、 Calibration Item &amp; Result :

## Alternating Current 60 Hz

Power Source Set Value					Standard Value(kW)	Indication Value(kW)	Error Value(%)	Uncertainty (%)
Mold/resistance	Phase/Wire	Voltage(V)	Current(A)	P.F.				
CH1	1P2W	220	1	1.0	0.22000	0.2201	+0.03	0.19
CH1	1P2W	220	5	1.0	1.10000	1.1005	+0.04	0.19
CH2	1P2W	220	1	1.0	0.22000	0.2200	0	0.08
CH2	1P2W	220	5	1.0	1.10000	1.0999	-0.01	0.19
CH3	1P2W	220	1	1.0	0.22000	0.2198	-0.08	0.19
CH3	1P2W	220	5	1.0	1.10000	1.0992	-0.08	0.19

The report issued by : Bang-yen Lai

1. The testing result is only responsible to the tested sample. The report can be fully duplicated only. To excerpt any part of this report is invalid unless permitted by TERTEC.
2. The contents on the report can not be used for advertisement, publication and merchandised activities.

60T-07-1801D

※ This document is invalid without the stamp.



財團法人

台灣大電力研究試驗中心  
Taiwan Electric Research & Testing Center

Electrical and Temperature Calibration Laboratory  
Report No. : EC20140157

## Calibration Report

### 二、Instruction :

1. Test method : according to Wattmeter calibration procedure, used of P,F source correction devices apply the watt/var Standard calculating the 1Phase Smart Power Meter error value(%).
2. Error(%)= $((E_{UUT}-E_{STD})/E_{STD}) \times 100\%$ ,  $E_{UUT}$  : 1Phase Smart Power Meter Energy Standard ,  $E_{STD}$  : Energy Standard
3. Expanded uncertainty the level of confidence is 95% and the coverage factor k=2.
4. Three phase compact smart Meter Energy used single phase two wire theorem.
5. The use of standard calibration devices traceable to national standards of weights and measurement laboratory and Pink Technology Co.Ltd (report No.E130546A, traceable data 2013.10.11, calibration cycle is one year).

60T-07-1803B

※ This document is invalid without the stamp.

Page 2 of 2

IMPARTIALITY SERVICE INNOVATION EFFICIENCY