



Three Phase Compact Smart Meter Calibration Report

Report number : EC2018008702

Issued date : 2018/06/04

Laboratory : Electrical and Temperature Calibration Laboratory

Address:No.6-6, Ronggong S Rd., Guanyin Township, Taoyuan
County 328, Taiwan, R.O.C.

Laboratory Accreditation Number : 0061



Taiwan Electric Research & Testing Center

Address:No.6-6, Ronggong S Rd., Guanyin Township, Taoyuan County 328, Taiwan, R.O.C.

TEL:886-3-483-9090

FAX:886-3-483-8119

E-mail:customer_service@ms.tertec.org.tw

Website:www.tertec.org.tw

- ◆ This report is calibration report.
- ◆ The testing result is only responsible to the tested sample.
- ◆ The report can be fully duplicated only. To excerpt any part of this report is invalid unless permitted by TERTEC.
- ◆ The contents on the report can not be used for advertisement, publication and merchandised activities.
- ◆ The report is invalid if without the seal on each page.
- ◆ Inquiry telephone:886-3-483-9090 ext. 8201



Calibration Report

Applicant : ICP DAS Co.,Ltd. Issued Date : 2018/06/04
 Address : No. 111, Guangfu N. Rd., Hukou Township, Calibration Date : 2018/05/24
 Hsinchu County 30351, Taiwan,(R.O.C.)
 Equipment : Power Meter Temperature : 23 ± 2°C
 Manufacturer : ICP DAS Humidity : 50 ± 10%
 Model No. : PM-3133-240 Procedure : 60I-07-1812,60I-07-1818
 Serial No. : PM3133CCR0kuJBR00022

Remark : 1. Calibration items with the external current transformer(CT1,CT2,CT3) which Manufacturing number are the same as serial No. of the equipment).
 2. The reading was obtained from the RS485 communication interface of the calibration product and the computer software(ICP DAS PM-3133 ver1.17) of the manufacturer.

Calibration Standard :

Equipment	Manufacturer / Model No.	I.D. Number	Cal. Source / Cal. Date/ Report No/Cycle
Three-Phase Watt/Var.	RADIAN/RD-30-211	300130	TERTEC / 107.03.16 / EC1070012 /1 year
Multifunction Standard	FLUKE/5500A	6670008	Pink Technology / 2017.08.15/P708046-C/1 year

一、 Calibration Item & Result :

1. AC Watt (60Hz)

Item	Setting			Standard(kW)	Reading(kW)	Error(%)	Uncertainty(%)
	Voltage(V)	Current (A)	P.F				
A Phase	220	5	1.0	1.1000	1.0973	-0.25	0.19
B Phase	220	5	1.0	1.1000	1.0992	-0.08	0.19
C Phase	220	5	1.0	1.1000	1.0990	-0.09	0.19

2. AC Voltage (60Hz)

Item	Standard (V)	Reading (V)	Error(%)	Uncertainty(%)
A Phase	100.0000	99.9938	-0.01	0.19
B Phase	100.0000	99.9910	-0.01	0.19
C Phase	100.0000	99.9885	-0.01	0.19

The Report Issued by : *Bang-yenn Lai*

1. The testing result is only responsible to the tested sample. The report can be fully duplicated only. To excerpt any part of this report is invalid unless permitted by TERTEC.
 2. The contents on the report can not be used for advertisement, publication and merchandised activities.



Calibration Report

二、Instruction :

1. Calibration method is refer to the laboratory Watt calibration instructions, It uses adjustable Multi-Function Standard to calculate error (%).
2. Calibration method is refer to the laboratory AC current calibration instructions, use stable AC power to source to calculate error (%).
3. $Error(\%) = ((EUUT - ESTD) / ESTD) \times 100\%$.
4. The level of confidence of Expanded Uncertainty is 95% and the coverage factor $k=2$.
5. The use of standard calibration devices is traced back to National Measurement Laboratory(Report No.. E170502A, traceable data 2017-08-29, calibration cycle is 1 year) and Pink Technology Co. Ltd (Report No. P708046-C, traceable data 2017-08-15, calibration cycle is 1 year).

